



DRUG REGULATORY AUTHORITY OF PAKISTAN

DIVISION OF QUALITY ASSURANCE & LABORATORY TESTING

MEDICAL PRODUCT ALERT

DRAP ALERT NO. N° I/S/06-26-30

MEDICAL DEVICES DECLARED SUBSTANDARD BY CENTRAL DRUG LABORATORY KARACHI

Date: 01st June, 2026

Target Audience:

- National Regulatory Field Force of DRAP and Provincial Drug Control Departments.
- Healthcare Professionals (Physicians, Pharmacists & Nurses).
- General Public.

Alert Summary:

Central Drugs Laboratory (CDL), Karachi has informed that samples of the below-mentioned Auto Disable Syringes have been declared “**Substandard**” due to failure of the **Self-Destructive Performance / Auto-Disability Test (re-use prevention feature)**.

S#	Product Name	Manufacturer / Importer & Address	Batch No.	Remarks
1	Zindagi Auto Disable Syringes 3ml/cc (MDMR-000117)	Manufacturer: M/s Al-Badar Manufacturing (Pvt.) Ltd. Plot No. 193/3, Road No. 7, Industrial Estate, Gadoon Amazai, District Swabi, Khyber Pakhtunkhwa, (ELM-0029)	3-01B-26	Substandard
2	Zindagi Auto Disable Syringes 3ml/cc (MDMR-000117)		3-01A-26	Substandard
3.	Biosoft Auto Lock Safety Syringe 3cc/ml (MDIR-0002506)	Manufacturer: M/s Jiangxi Fenglin Medical Appliances Co. Ltd, Jiangxi China Importer: M/s Hashir Surgical Services Office No 16, Street 1, F-2, Phase 6, Hayatabad, Peshawar	20250810	Substandard
4.	I.CO Safety Syringe 5ml (MDIR-0002291)	Importer: M/s Capra-Med (Pvt) Ltd First Floor 81, Broadway Commercial, Paragon City Barki Road, Lahore Manufacturer: M/s Med EL Dawlaico International Company for Medical Necessities	2510	Substandard
5.	Injekt Auto Disable Syringe 3ml (MDMR-000147)	Manufacturer: Lahore Medical Instruments (Pvt.) Ltd. Kasur	3-018	Substandard



DRUG REGULATORY AUTHORITY OF PAKISTAN

DIVISION OF QUALITY ASSURANCE & LABORATORY TESTING

6.	Yashfaeen Sterile Self Destruction Safety Syringe 3ml [Auto Disable Syringe] (MDIR-0006365)	Importer: M/s Tech Zone Ground Floor, Weal House 8, Faiz Road, Old Muslim Town, Lahore. Manufacturer: M/s Shandong Caremed Medical Products Co., Ltd., (Yantai) Industrial park, Dingtao County, Heze 274100 Shandong PR China	20250930	Substandard
7.	Yashfaeen Sterile Self Destruction Safety Syringe 5ml [Auto Disable Syringe] (MDIR-0006365)	Importer: M/s Tech Zone Ground Floor, Weal House 8, Faiz Road, Old Muslim Town, Lahore. Manufacturer: M/s Shandong Caremed Medical Products Co., Ltd., (Yantai) Industrial park, Dingtao County, Heze 274100 Shandong PR China	20250930	Substandard

Risk Statement:

Auto-disable and auto-destruct syringes are designed to become permanently inoperable after a single use. Failure of the self-destructive mechanism may permit reuse of the device, increasing the risk of transmission of blood-borne pathogens including Hepatitis B Virus (HBV), Hepatitis C Virus (HCV) and Human Immunodeficiency Virus (HIV). Reuse of such defective syringes may also result in cross-contamination, healthcare-associated infections and serious public health consequences, particularly during immunization programmes and other invasive medical procedures.

Action initiated: -

The field force of DRAP and Provincial Drug Control Departments has been directed to immediately conduct market surveys for detection of presence and removal of the affected batches from the market. Manufacturers, importers and distributors have been directed to initiate recall of the affected batches in accordance with applicable regulatory requirements.

Advice for Pharmacies/Medical stores: -

All pharmacists and chemists working at distributions and pharmacies should immediately check their stocks and stop supplying the above-mentioned batches. The remaining stocks should be quarantined and returned to the supplier/company. Regulatory field force of all federating units should also increase surveillance in the market to ensure effective recall of the affected products.

Advice for Healthcare Professionals: -

DRAP requests increased vigilance within the supply chains of institutions/pharmacies/healthcare facilities likely to be affected by batches of mentioned products. Adverse reactions or quality problems experienced with the use of above



DRUG REGULATORY AUTHORITY OF PAKISTAN

DIVISION OF QUALITY ASSURANCE & LABORATORY TESTING

mentioned batches of the drug product are to be reported to the National Pharmacovigilance Centre (NPC), DRAP using Adverse

Event Reporting Form or online through this [link](#). Further information of reporting problems to DRAP is available on this [link](#).

Advice for Consumers/general public: -

Consumers and healthcare facilities in possession of the affected batches should immediately stop using these products and contact the supplier, distributor or manufacturer for return or replacement. Any adverse event or quality issue associated with the use of these products should be reported to DRAP/National Pharmacovigilance Centre.



Drug Regulatory Authority of Pakistan

محفوظ، موثر اور معیاری اشیائے علاج



DRAP, Islamabad



+92 051 9255969



gsms@dra.gov.pk